



**REQUEST FOR QUOTATION**

Date: May 2, 2024  
 RFQ.: BSVP-ILO 2024-021

Name of Company: \_\_\_\_\_  
 Address: \_\_\_\_\_  
 Business Permit No.: \_\_\_\_\_  
 TIN No.: \_\_\_\_\_  
 PhilGEPS Registration No.: \_\_\_\_\_

**PROCUREMENT: Supply and Delivery of Multi-Parameter Measuring Tool for Waste Water Monitoring for Iloilo International Airport**

Please quote your best offer for the item/s described below, subject to the Terms and Conditions provided on this request for quotation. Submit your quotation duly signed by you or your duly representative and must send copies of eligibility requirements (**Business Permit, Omnibus Sworn Statement, PhilGEPS Certificate of Registration, and Special Power of Attorney for Sole Proprietorship/ Secretary Certificate for Corporation and Cooperative, if applicable**) not later than **May 9, 2024 at 9:00 AM**.

Interested bidder must submit this RFQ and Price Quotation in the Bidder's/Supplier's Letterhead sealed in an envelope addressed to the BAC Chairman. Kindly review Terms and Conditions attached herewith.

**(SGD.) ROBERTO B. MONTELLJAO JR.**

Procurement Officer

**TERMS AND CONDITIONS:**

1. Bidders must provide **correct and accurate** information required in this form.
2. Only **authorized representative** shall be allowed to sign the Request For Quotation in behalf of the owner/ corporation.
3. Bidders must quote for all the items.
4. Price quotation/s must be valid for a period of Thirty (30) calendar days from the date of submission.
5. Price quotation/s, to be denominated in Philippine peso shall include all taxes, duties and/or levies payable.
6. Quotations exceeding the Approved Budget for the Contract shall be rejected.
7. Award of contract shall be made to the lowest quotation (for goods and infrastructure) or, the highest rated offer (for consulting services) which complies with the minimum technical specifications and other terms and conditions stated herein.
8. Any interlineations, erasures or overwriting shall be valid only if they are signed or initiated by you or any of your duly authorized representative's.
9. The item/s shall be delivered within **Ninety (90)** Calendar days from receipt of Purchase Order.
10. The GPPB-TSO shall have the right to inspect and/or to test the goods to confirm their conformity to the technical specifications.
11. Liquidated damages equivalent to one tenth of one percent (0.001%) of the value of the goods not delivered within the prescribed delivery period shall be imposed per day of delay. The GPPB-TSO shall rescind the contract once the cumulative amount of liquidated damages reaches ten percent (10%) of the amount of the contract, without prejudice to other courses of action and remedies open to it.

After having carefully read and accepted the Terms and Conditions, I/we submit our quotation/s for the item/s as follows:

ITEM DESCRIPTION	Quantity (QTY)		Approved Budget of the Contract	OFFER						
				PRICE			BRAND indicate brand or generic (mandatory)	Compliance with Technical Specifications (Please Check)		REMARKS
				QTY	Unit Price	Total Price		Yes	No	
			<b>PHP 373,137.56</b>							
1.0	Multiparameter Analyzer (see attached specs)		1	set						
	Consumables									
2.0	COD Reagent (100 samples/box)		1	box						
3.0	Ammonia Nitrogen Reagents (100 samples/box)		1	box						
4.0	Total Phosphorus Reagents (100 samples/box)		1	box						
5.0	Total Nitrogen Reagents (100 samples/box)		1	box						
6.0	Tube Rack		1	pc						
7.0	Pipette 1ml		1	pc						
8.0	Pipette 5ml		1	pc						

ITEM DESCRIPTION	Quantity (QTY)	Approved Budget of the Contract	OFFER						
			PRICE			BRAND indicate brand or generic (mandatory)	Compliance with Technical Specifications (Please Check)		REMARKS
			QTY	Unit Price	Total Price		Yes	No	
		<b>PHP 373,137.56</b>							
9.0	<b>Portable PH Meter (see attached specs)</b>	1	set						
	<b>Consumables</b>								
10.0	Universal Adaptor 9V	1	unit						
11.0	PH Electrode	1	pc						
12.0	PH4 Buffer Solution 500 ml	1	bottles						
13.0	PH7 Buffer Solution 500 ml	1	bottles						
14.0	PH10 Buffer Solution 500ml	1	bottles						
15.0	Deionized Water 500ml	1	bottles						
16.0	<b>Dissolved Oxygen Meter Kit (see attached specs)</b>	1	set						
	<b>Consumables</b>								
17.0	DO Electrode Galvanic w/ ATC	1	pc						
18.0	DO Electrolyte Solution NAOH (30ml)	1	bottle						
	<b>TOTAL</b>								

## Specifications

### I. Multiparameter Analyzer

Measurements	COD (-15000mg/L) Ammonia (0-200mg/L) Total Phosphorous (10-100mg/L) Total Nitrogen (0-15mg/L) Turbidity, Color, Suspendid Solid, Organic, Inorganic, Metal, Pollutants
Curve number	180 pcs
Data Storage	40 thousands sets
Accuracy	COD≤50mg/L, ≤±8%; Cod>50mg/L, ≤±5%; TP≤±8%; other indicator ≤10
Repeatability	3%
Colorimetric method	By 16mm/25mm round tube
Resolution ratio	0.001Abs
Display screen	3.5" colorful LCD display screen
Battery capacity	Lithium battery 3.7V 3000mAh
Charging method	5W USB-Type C
Printer	External Bluetooth Printer
Host weight	0.6kg
Host size	224 x (108x78)mm
Instrument Power	0.5W
Ambient Temperature	40° C
Ambient Humidity	≤85%RH (No condensation)
Warranty	1 Year

### II. Portable PH Meter Kit

Kit Inclusions	pH meter and sensor, pH 4, pH 7 and pH 10 calibration solution, calibration certificate, 3 pcs AAA batteries and rugged carrying case
Datalogger	Auto: and Manual.
Memory Card	SD memory card 8G
Data Hold	Yes
Meter Dimension	175mm x 58mm x 32mm (with BNC connector)
Power Supply	AAA batteries (3 pcs) / 9V AC/DC Adaptor
Temperature	0~90° C
Temp. accuracy	±0.5° C
Measurement range	pH 0 to 14 pH
Accuracy	±0.02 pH + 2 digits

ITEM DESCRIPTION	Quantity (QTY)	Approved Budget of the Contract	OFFER					REMARKS	
			PRICE			BRAND indicate brand or generic (mandatory)	Compliance with Technical Specifications (Please Check)		
			QTY	Unit Price	Total Price		Yes		No
		<b>PHP 373,137.56</b>							

pH calibration                      pH 4, 7, 10, 3 points calibration  
Resolution                            0.01(pH)  
Automatic Temp                      Yes  
Dimension                            12 x 120mm  
Electrode body                      PC  
Sensor Type                         Glass Bulb  
ATC Temperature                    3.5  
Cable length                         1 meter  
Warranty                              1 Year

**II. Portable Dissolved Oxygen Kit**

Kit Inclusions                        DO meter, 3 pcs AAA batteries, 1 pc Electrode (DO galvanic type)  
Blacl carrying case, Electrolyte ( 0.5M NaOH), 1 pc Membrane cap,  
10 pcs Membrane, Red sandpaper ( for polishing DO electrode), 8GB SD  
card and calibration certificate  
Data Hold                             Yes  
Meter dimension                    175mm x 58mm x 32 mm (with BNC connector)  
Power Supply                        3 pcs AAA battery / 9V AC/DC  
Parameter                            DO, temperature  
SD Sampling time setting range    Auto: (0,2,5,10,15,30,60,120,300,600,900,1800 sec and 1 hr)  
and Manual.

**DO Electrode Specification**

Temperature                         0~90° C  
Temp. accuracy                     ±0.5° C  
Measurement range                0~199.99% (in saturation); 0.0~20 mg/L  
Accuracy                            ±2% if full scale + 1 digit  
Resolution                         0.1%, 0.1 mg/L  
Calibration                         100% Air-saturated  
Flow condition                     0.3 mL/s  
Dimension                         12x120mm  
Electrode body                     ABS  
Sensor Type                        Galvanic  
ATC Temp. sensor probe port    3.5Ø mm phone jack (10K Ω resistance)  
Cable length                        3 meters

---

*Signature over Printed Name*

---

Contact Number (Landline and/ or Cellphone Nos)/Email Address