



Republic of the Philippines
**CIVIL AVIATION AUTHORITY OF THE
PHILIPPINES**

SUPPLEMENTAL/BID BULLETIN NO. 3

**PURCHASE/INSTALLATION OF METEOROLOGICAL INSTRUMENTS FOR VARIOUS
AIRPORTS (SINGLE SITE)**

(BID NO. 24-080-10-BRAVO)

Date: ~~NOV 08 2024~~

Please note the following amendments/revisions/clarifications to the Bidding Documents:


Reference	Remarks	Requirement/s
I. TECHNICAL SPECIFICATIONS		TECHNICAL SPECIFICATIONS
B.1.3.1.5. The contractor shall supply a digital barometric pressure sensor that shall be installed either at the runway inside the AWS cabinet or Cabroom.	Amendment of BB No. 2	B.1.3.1.5. The contractor shall supply a digital barometric pressure sensor that shall be installed either at the sensor site inside the AWS cabinet or FSS Bldg/ATC Tower.
B.1.4.1.4. The supplied aviation weather display shall be portable and be installed directly or fit to the ATC console at ATC cab room/FSS Building.	Revision of Technical Specification	B.1.4.1.4. The supplied aviation weather displays shall be portable and be installed directly or fit to the ATC console at ATC cab room/FSS Building, laptop or PC is not considered as portable MET display for this project. The weather displays are data panel display and wind panel display. Both comes with built-in visual and audible alarms and can be mounted on a standard IEC (<i>International Electrotechnical Commission</i>) panel or its equivalent in the Philippines.
2. Query		
Is a consortium with a foreign company allowed in lieu of joint venture for this tender?	Clarifications	<p>Under this Goods procurement project:</p> <p><i>"Bidding is restricted to Filipino citizens/sole proprietorships, partnerships, or organizations with at least sixty percent (60%) interest or outstanding capital stock belonging to citizens of the Philippines, and to citizens or organizations of a country the laws or regulations of which grant similar rights or privileges to Filipino citizens, pursuant to RA 5183."</i></p> <p><i>Also, refer to 23.4.1.1 (e) of the 2016 revised IRR of RA No. 9184.</i></p>


<p>We would like to request for a 6-week extension from the date of submission of bids to have ample time to inspect the 4 airport sites and prepare all the documentary requirements, drawings and design for bid submission since we will be having a consortium with a foreign company. Translation of required documents and PhilGeps registration will take several weeks.</p>		<p>Please refer to BB No. 2 for the new schedule of submission and opening of bids.</p>
---	--	---

This bid bulletin shall be an integral part of the Bidding Documents and the same shall be enclosed in the technical bid envelop/component and shall be marked/tapped accordingly.

For the information and guidance of all concerned:

Approved by:


GARY M. JADIE
 End-User


ATTY. DANJUN G. LUCAS
 Chairperson, Bids and Awards Committee (Bravo)